

DIGITAL MANOMETER

Product Codes: 06104

Precision, handheld, digital pressure reference manometer conveniently calibrates pressure transducers. Can be recalibrated in the field for zero, span, and linearity.

- Precision, handheld, digital pressure reference manometer conveniently calibrates pressure transducers. Can be recalibrated in the field for zero, span, and linearity.
- Accurate primary standards to calibrate devices locally
- 1 point or 5 point calibration options
- Engineering units - Inches of H₂O (select 4°C, 20°C, or 60°C), PSI, mm of Hg, inches of Hg, mBars, Bars, kPa, kg/cm², cm of H₂O and user selectable scaling.
- Accuracy: ±0.05% FS
- Media Compatibility – DN: Non-isolated differential sensor for clean, dry, non-corrosive gases. GI & AI: Isolated Gauge or Absolute sensor for fluids compatible with 316L SS



No part of this publication may be reproduced, stored in a retrieval system or transmitted in any form or by any means, electronic, mechanical, photocopying, recording, scanning or otherwise without the prior written permission of the Publisher.

Limit of Liability/Disclaimer of Warranty: While the publisher and author have used their best efforts in preparing this document, they make no representation or warranties with respect to the accuracy or completeness of the contents of this document and specifically disclaim any implied warranties of merchantability or fitness for a particular purpose. No warranty may be created or extended by sales representatives or written sales material. The advice and strategies contained herein may not be suitable for your situation. Neither the Publisher nor author shall be liable for any loss of profit or any other commercial damages, including but not limited to special, incidental, consequential or other damages.

455 Boleskine Road, Victoria B.C. Canada, V8Z 1E7
info@ViVITroLabs.com | (250) 388-3531 | www.ViVITroLabs.com

Copyright © 2016 by ViVITro Labs Inc. All rights reserved.

ViVITRO Labs Inc.
First in Cardiovascular Device Testing